

Certificate of Conformity

Number:

C251-0074/18

Project file:

C20180408

Product:

UVC Room Disinfection Device

Type reference:

STRPR-D

Trademark:

SteriPro UVC disinfection technology

Applicant:

UVC Solutions d.o.o.

Ljubljanska cesta 20, SI-3000 Celje, Slovenia

Manufacturer:

UVC Solutions d.o.o.

Ljubljanska cesta 20, SI-3000 Celje, Slovenia

Place of manufacture: UVC Solutions d.o.o.

Ljubljanska cesta 20, SI-3000 Celje, Slovenia

This certificate is granted subject to the SIQ's rules on product certification. SIQ certifies the conformity of the products with the requirements of the listed standards.

Ratings:

120-240 V~; 50-60 Hz; PRATED = 2,3 kW

Maximum clock frequency: 2,4 GHz

Protection class: I

Standard:

EN 61326-1:2013 EN 61000-3-2:2014

EN 61000-3-3:2013

Test report:

T251-0407/18

Remarks:

This certificate shall apply to the products identical to the tested sample and

shall remain valid for the period of 3 years until 2021-06-21 or until the

validity date of the listed standards, whichever occurs earlier.

Product stated is also in conformity with Draft EN 303 446-1 V1.1.0.

Date:

2018-06-21

Authorized signature:

Bojan Pečavar

Only integral publication of this certificate is allowed. This certificate may only be reproduced in its entirety and without any changes. On request SIQ will give information about the validity of the certificate.